	Application No.	Applicant(s)
Notice of Allowability	09/926,016	MURGATROYD ET AL.
	Examiner	Art Unit
	Amy R Cohen	2859
Th MAILING DATE of this communication appe All claims being allowable, PROSECUTION ON THE MERITS IS herewith (or previously mailed), a Notice of Allowance (PTOL-85) NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RI- of the Office or upon petition by the applicant. See 37 CFR 1.313	(OR REMAINS) CLOSED in this apported or other appropriate communication GHTS. This application is subject to	plication. If not included will be mailed in due course. THIS
1. 🖾 This communication is responsive to <u>Amendment filed 1/23</u>	<u>3/04</u> .	
2. The allowed claim(s) is/are <u>1-87</u> .		
3. The drawings filed on 23 January 2004 are accepted by the Examiner.		
 4. Acknowledgment is made of a claim for foreign priority un a) All b) Some* c) None of the: 1. Certified copies of the priority documents have 2. Certified copies of the priority documents have 3. Copies of the certified copies of the priority documents have International Bureau (PCT Rule 17.2(a)). * Certified copies not received: 	been received. been received in Application No	
Applicant has THREE MONTHS FROM THE "MAILING DATE" on noted below. Failure to timely comply will result in ABANDONM THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.		complying with the requirements
5. A SUBSTITUTE OATH OR DECLARATION must be submit INFORMAL PATENT APPLICATION (PTO-152) which give		
 6. CORRECTED DRAWINGS (as "replacement sheets") must be submitted. (a) including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached 1) hereto or 2) to Paper No./Mail Date (b) including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d). 7. DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL. 		
 Attachm nt(s) 1. ☑ Notice of References Cited (PTO-892) 2. ☐ Notice of Draftperson's Patent Drawing Review (PTO-948) 3. ☐ Information Disclosure Statements (PTO-1449 or PTO/SB/0 Paper No./Mail Date	6. ☐ Interview Summary Paper No./Mail Da 8), 7. ☐ Examiner's Amendr	tè

Application/Control Number: 09/926,016

Art Unit: 2859

DETAILED ACTION

Allowable Subject Matter

1. Claims 1-87 are allowed.

Reasons for Allowance

2. The following is an examiner's statement of reasons for allowance:

Claims 1-86: The prior art of record does not disclose or suggest an optical spectrum analyzer comprising: means to apply a variable axial force to the first section of fibre, to thereby tune the peak wavelength of the grating over a desired wavelength range in combination with the remaining limitations of the claims.

Claim 87: The prior art of record does not disclose or suggest a tunable optical filter comprising: means to apply a variable axial force to the first section of fibre, to thereby tune the peak wavelength of the grating over a desired wavelength range in combination with the remaining limitations of the claims.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Conclusion

3. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. The following patents disclose optical spectrum analyzers Boettcher et al. (U. S.

Page 2

Application/Control Number: 09/926,016

Art Unit: 2859

Patent No. 6,738,536), Kempen et al. (U. S. Patent No. 6,696,682), Valente et al. (U. S. Patent No. 6,658,171), Satorius (U. S. Patent No. 6,647,159), Lieberman et al. (U. S. Patent No.

6,650,810), and Engel et al. (U. S. Patent No. 6,594,081).

4. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Amy R Cohen whose telephone number is (571) 272-2238. The examiner can normally be reached on 8 am - 5 pm, M-F.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Diego F. Gutierrez can be reached on (571) 272-2245. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

ARC May 28, 2004

> Diego Gutierrez Supervisory Examiner Tech Center 2800

Page 3